

تقديم

الهيئة جهة مستقلة الغرض الأساسي لها هو القيام بتنظيم ومراقبة الغذاء والدواء والأجهزة والمنتجات الطبية، ومن مهامها وضع اللوائح الفنية والمواصفات في مجالات الغذاء والدواء والأجهزة والمنتجات الطبية سواء كانت مستوردة أو مصنعة محلياً، وقد قام قطاع الأجهزة والمنتجات الطبية بالهيئة ضمن برنامج عمل الفريق رقم (SFDA/MDS/TC 62) "فريق عمل مواصفات الأجهزة الكهربائية في التطبيقات الطبية" بتبني المواصفة الدولية رقم (IEC 60601-2-19: 2016) "أجهزة الكهربية الطبية-الجزء 2-19: المتطلبات الخاصة بسلامة حاضنات الأطفال"، والتي أصدرتها "المنظمة الدولية الكهروتقنية" وذلك بلغتها الأصلية. وقد اعتمدت هذه المواصفة كمواصفة سعودية متبناة بالمطابقة بلغتها الأصلية وقد تم إقرار تبني المواصفة من معالي الرئيس التنفيذي للهيئة بقرار رقم (.....) وتاريخ

Foreword

The Saudi Food and Drug Authority (SFDA) is an independent organization mainly responsible for regulating imported/local food, drug and medical devices which includes, inter alia, setting their standards. International Standard No. (IEC 60601-2-19: 2016) "Medical electrical equipment PART 2-19 : Particular Requirement for the Safety of BABY INCUBATORS" issued by "International electrotechnical Commission" has been adopted identically in its original language. This standard is adopted with modifications in its original language as a national standard and approved by SFDA CEO decision No (...) on (date)

Scope:

Replacement:

This International Standard applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of INFANT INCUBATORS, as defined in 201.3.209 of this standard, also referred to as ME EQUIPMENT.

If a clause or subclause is specifically intended to be applicable to ME EQUIPMENT only, or to ME SYSTEMS only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to ME EQUIPMENT and to ME SYSTEMS, as relevant.

HAZARDS inherent in the intended physiological function of ME EQUIPMENT or ME SYSTEMS within the scope of this standard are not covered by specific requirements in this standard except in 7.2.13 and 8.4.1 of the general standard.

NOTE See also 4.2 of the general standard.

This particular standard specifies safety requirements for INFANT INCUBATORS but alternate methods of compliance with a specific clause by demonstrating equivalent safety will not be judged as non-compliant if the MANUFACTURER has demonstrated in his RISK MANAGEMENT FILE that the RISK presented by the HAZARD has been found to be of an acceptable level when weighed against the benefit of treatment from the device.

This particular standard does not apply to:

- devices supplying heat via BLANKETS, PADS or MATTRESSES in medical use; for information see IEC 80601-2-35 [3]2);
- INFANT RADIANT WARMERS; for information, see IEC 60601-2-21 [2];
- INFANT TRANSPORT INCUBATORS, for information, see IEC 60601-2-20 [1];
- INFANT PHOTOTHERAPY EQUIPMENT, for information see IEC 60601-2-50 [4].