

Safety Communication

رسالة سلامة

Black dirt found in water of Hydrox

Device/ Product Description:	Disposable Prefilled Humidifier	
Brand:	UWC Hydrox Prefilled Humidifier	
Part Number:	<u>Lot Number</u>	LN-08-80044
	<u>Catalogue Number</u>	HA65-000
Manufacturer:	UWHM SDN. BHD.	
Problem:	Black dust particle contaminants in the water of the product.	
Recommendation/Actions:	<ol style="list-style-type: none">1. Review this notice and ensure that affected personnel are aware of the contents.2. inspect the inventory to identify the affected products of the above mentioned Lot & Catalogue numbers3. Contact the authorized representative for replacement. <p>If you or your patient have a problem with a medical device, please report the problem to SFDA through: NCMDR Vigilance system 19999 unified call center</p>	

Devices/Products photo:		
Authorized Representative Details	AR name:	KAL Medical Development Foundation
	Assigned Contact Person:	Eng. Mounir Halwani
	Mobile/Phone:	0567373073 Tel: +966(11)2176444 & +966(11)2012960
	Email:	mounir@kalmedical.com.sa mugbel@kalmedical.com.sa osama@kalmedical.com.sa

