

Safety Communication

رسالة سلامة

May Have Cybersecurity Vulnerabilities

Device/ Product Description:	Interventional Workstation Software	
Manufacturer:	Philips Healthcare	
Problem Summary:	<p>A vulnerability allow attacker within the hospital network to shut down or restart the workstation.</p> <p>For more information, please check Here.</p> <p>If you think you had a problem with your device or a device your patient uses, please report the problem to SFDA through: NCMDR Vigilance system 19999 unified call center</p>	
	AR name:	Philips Healthcare Saudi Arabia Ltd.
	Assigned Contact Person:	Mohammed Alsamhan
	Mobile/Phone:	+966 503879145
	Email:	mohammed.alsamhan@philips.com