

Safety Communication

رسالة سلامة

Deviations of sterilization process parameters

Device/ Product Description:	Prismaflex Sets	
Affected product:	Check the attached “Field Safety Notice” Here .	
Manufacturer:	Baxter Healthcare	
Problem:	<p>The above mentioned product are supplied to the market in sterile status, following the Ethylene Oxide sterilization process performed overtime by sterilization service providers (Steril Milano Srl).</p> <p>Possible deviations of sterilization process parameters (in specific Lots/Batches) required the legal manufacturer to conduct a Field Safety Corrective Action (FSCA).</p>	
Recommendation /Actions:	<ul style="list-style-type: none"> • Make sure that this document is reached to the end-users. • Go through the links above to find out all the details and instructions. • Identify and quarantine any affected products. • Contact the authorized representative for the required support. <p>If you think you had a problem with your device or a device your patient uses, please report the problem to SFDA through: NCMDR Vigilance system (19999) unified call center</p>	
	AR name:	Baxter AG

Authorized Representative Details	Assigned Contact Person:	Ahmed Alhamad
	Mobile/Phone:	+(966) 554428638
	Email:	ahmed_alhamad@baxter.com

