

Safety Communication

رسالة سلامة

May Be Susceptible to PrintNightmare Vulnerability

| | | |
|--|---|-------------------------------------|
| Device/ Product Description: | Various Clinical laboratory Analyzers | |
| Manufacturer: | Beckman Coulter Inc. | |
| Problem: | <p>The manufacturer states that it is evaluating the risk of the PrintNightmare vulnerability on its products. The manufacturer also states that where appropriate, teams are validating the installation of the Microsoft security patch and developing specific customer instructions for those systems.</p> <p>For more information, please check Here.</p> <p>If you think you had a problem with your device or a device your patient uses, please report the problem to SFDA through: NCMDR Vigilance system 19999 unified call center</p> | |
| Authorized Representative Details | AR name: | Beckman Coulter Saudi Arabia Co Ltd |
| | Assigned Contact Person: | Saeed Alqahtani |
| | Mobile/Phone: | +(966) 505404792 |
| | Email: | ssalqahtani@beckman.com |