

المملكة الصربية السعودية Saudi Food & Drug Authority تامملكة الضفاء والدواء

Medical Device Sector Surveillance & Biometrics Executive Department قطاع الأجهزة والمنتجات الطبية الإدارة التنفيذية للرقابة والقياسات الحيوية

Safety Communication

رسالة سلامة

Possible systems security vulnerabilities

Device/ Product Description:	Various Systems Using Apache Java Log4j Library Utility
Affected product:	Manufacturer NameLinkPhilipsPhilipsSiemens HealthineersSiemensVarian Medical SystemsVarianBeckman CoulterBeckman
Problem:	Apache has confirmed that a critical Remote Code Execution vulnerability (CVE-2021-44228 and CVE-2021-45046) exists in their Log4j utility. Log4j is an open source, java-based logging utility.
Recommendation /Actions:	For more information and support please contact your supplier. Also, you can visit Apache Log4j Security Vulnerabilities via: Apache and the following websites: Philips, Siemens, Varian, Beckman If you think you had a problem with your device or a device your patient uses, please report the problem to SFDA through: NCMDR Vigilance system (19999) unified call center

SG-2112-373-H 01/23/2022